

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	3.991 °C	4.4 °C	0.409 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.222 °C	15.6 °C	0.378 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.163 °C	30.2 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	27.5 %HR	-2.866 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.100 %HR	48.1 %HR	-2.000 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.294 %HR	61.9 %HR	-3.394 %HR	± 3.0 %HR	± 1.9E+00 %HR