

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.163 °C	30.2 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	27.5 %HR	-2.866 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.100 %HR	48.1 %HR	-2.000 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	66.9 %HR	-3.343 %HR	± 3.0 %HR	± 1.9E+00 %HR