

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
9.8 RPM	10 RPM	0.2 RPM	-	±	5.8E-01 RPM
33.0 RPM	33 RPM	0.0 RPM	-	±	5.8E-01 RPM
64.2 RPM	64 RPM	-0.2 RPM	-	±	5.8E-01 RPM