

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.1	0.09942	0.580	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.19942	0.580	± 0.6	± 4.3E-01 mm					
0 a 0.3	0.29945	0.550	± 0.7	± 4.3E-01 mm					
0 a 0.4	0.39941	0.590	± 0.7	± 4.3E-01 mm					
0 a 0.5	0.49952	0.480	± 0.8	± 4.3E-01 mm					
0 a 0.6	0.59937	0.630	± 0.8	± 4.3E-01 mm					
0 a 0.7	0.69936	0.640	± 0.8	± 4.3E-01 mm					
0 a 0.8	0.79956	0.440	± 0.9	± 4.3E-01 mm					
0 a 0.9	0.89946	0.540	± 0.9	± 4.3E-01 mm					
0 a 1	0.99939	0.610	± 1	± 4.3E-01 mm					
0 a 2	2.00018	0.530	± 1.4	± 4.8E-01 mm					
0 a 3	3.00028	0.380	± 1.8	± 4.8E-01 mm					
0 a 4	4.00027	0.350	± 2.2	± 4.8E-01 mm					
0 a 5	5.00020	-0.199	± 2.6	± 4.8E-01 mm					
0 a 6	6.00021	-0.210	± 3	± 3.6E-01 mm					
0 a 7	7.00024	-0.241	± 3.4	± 3.6E-01 mm					
0 a 8	8.00025	-0.248	± 3.8	± 3.6E-01 mm					