

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	4.041 °C	3.9 °C	-0.141 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	14.9 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.3 °C	0.238 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.458 %HR	27.5 %HR	-2.958 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.294 %HR	51.8 %HR	1.506 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.079 %HR	67.8 %HR	2.721 %HR	± 3.0 %HR	± 1.9E+00 %HR