

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 4 °C              | 4.041 °C          | 3.8 °C    | -0.241 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 15 °C             | 15.021 °C         | 14.8 °C   | -0.221 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.062 °C         | 30.3 °C   | 0.238 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.458 %HR        | 28.6 %HR  | -1.858 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.294 %HR        | 51.7 %HR  | 1.406 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 65 %HR            | 65.079 %HR        | 63.9 %HR  | -1.179 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |