

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.408 °C	10.6 °C	0.192 °C	± 1.0 °C	± 4.0E-01 °C
15 °C	15.522 °C	15.6 °C	0.078 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.1 °C	0.065 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.7 °C	0.250 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.363 °C	29.9 °C	-0.463 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	25.253 %HR	28.1 %HR	2.847 %HR	± 10.0 %HR	± 1.9E+00 %HR
30 %HR	30.468 %HR	29.0 %HR	-1.468 %HR	± 10.0 %HR	± 1.9E+00 %HR
40 %HR	40.284 %HR	40.2 %HR	-0.084 %HR	± 10.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	48.1 %HR	-1.969 %HR	± 10.0 %HR	± 1.9E+00 %HR
60 %HR	60.591 %HR	63.5 %HR	2.909 %HR	± 10.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	72.8 %HR	2.455 %HR	± 10.0 %HR	± 1.9E+00 %HR