

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.408 °C	10.3 °C	-0.108 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.522 °C	15.3 °C	-0.222 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.0 °C	-0.035 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.4 °C	-0.050 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.1 °C	-0.263 °C	± 0.5 °C	± 4.0E-01 °C
25 %HR	25.253 %HR	27.7 %HR	2.447 %HR	± 3.0 %HR	± 1.9E+00 %HR
30 %HR	30.468 %HR	27.3 %HR	-3.168 %HR	± 3.0 %HR	± 1.9E+00 %HR
40 %HR	40.284 %HR	37.1 %HR	-3.184 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	47.6 %HR	-2.469 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.591 %HR	62.1 %HR	1.509 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	72.9 %HR	2.555 %HR	± 3.0 %HR	± 1.9E+00 %HR