

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 10 °C                      | 10.408 °C                  | 10.6 °C            | 0.192 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 15 °C                      | 15.522 °C                  | 15.8 °C            | 0.278 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.035 °C                  | 20.0 °C            | -0.035 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.450 °C                  | 25.7 °C            | 0.250 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 30.363 °C                  | 30.5 °C            | 0.137 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 %HR                     | 25.253 %HR                 | 25.5 %HR           | 0.247 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 30 %HR                     | 30.468 %HR                 | 29.9 %HR           | -0.568 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 40 %HR                     | 40.284 %HR                 | 38.7 %HR           | -1.584 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.069 %HR                 | 53.0 %HR           | 2.931 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 60 %HR                     | 60.591 %HR                 | 61.7 %HR           | 1.109 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.345 %HR                 | 73.0 %HR           | 2.655 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |