

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.408 °C	10.6 °C	0.192 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.522 °C	15.8 °C	0.278 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.0 °C	-0.035 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.7 °C	0.250 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.5 °C	0.137 °C	± 0.5 °C	± 4.0E-01 °C
25 %HR	25.253 %HR	25.5 %HR	0.247 %HR	± 3.0 %HR	± 1.9E+00 %HR
30 %HR	30.468 %HR	29.9 %HR	-0.568 %HR	± 3.0 %HR	± 1.9E+00 %HR
40 %HR	40.284 %HR	38.7 %HR	-1.584 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	53.0 %HR	2.931 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.591 %HR	61.7 %HR	1.109 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	73.0 %HR	2.655 %HR	± 3.0 %HR	± 1.9E+00 %HR