

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.408 °C	10.5 °C	0.092 °C	± 1.0 °C	± 4.0E-01 °C
15 °C	15.522 °C	15.5 °C	-0.022 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.1 °C	0.065 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.4 °C	-0.050 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.3 °C	-0.063 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	25.253 %HR	28.2 %HR	2.947 %HR	± 10.0 %HR	± 1.9E+00 %HR
30 %HR	30.468 %HR	28.7 %HR	-1.768 %HR	± 10.0 %HR	± 1.9E+00 %HR
40 %HR	40.284 %HR	38.0 %HR	-2.284 %HR	± 10.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	51.3 %HR	1.231 %HR	± 10.0 %HR	± 1.9E+00 %HR
60 %HR	60.591 %HR	60.1 %HR	-0.491 %HR	± 10.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	73.1 %HR	2.755 %HR	± 10.0 %HR	± 1.9E+00 %HR