

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.862 °C	29.9 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.5 %HR	-0.070 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.407 %HR	50.1 %HR	-0.307 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	70.5 %HR	-0.050 %HR	± 3.0 %HR	± 1.9E+00 %HR