

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.071 °C	15.6 °C	0.569 °C	± 0.5 °C	± 4.0E-01 °C
# 15 °C	14.780 °C	15.4 °C	0.600 °C	± 0.5 °C	± 7.4E-02 °C
# 35 °C	34.795 °C	35.0 °C	0.235 °C	± 0.5 °C	± 7.4E-02 °C
# 45 %HR	41.848 %HR	44.5 %HR	2.672 %HR	± 0.5 %HR	± 7.4E-02 %HR
70 %HR	69.374 %HR	70.7 %HR	1.366 %HR	± 3 %HR	± 1.9E+00 %HR