

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.514 °C	12.1 °C	-0.414 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.555 °C	27.4 °C	-0.155 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.296 °C	42.5 °C	0.204 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.656 %HR	50.8 %HR	-1.856 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.084 %HR	69.5 %HR	-2.584 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.125 %HR	92.4 %HR	0.275 %HR	± 3.0 %HR	± 1.9E+00 %HR