

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.514 °C	12.7 °C	0.186 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.555 °C	27.4 °C	-0.155 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.296 °C	42.7 °C	0.404 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.656 %HR	51.2 %HR	-1.456 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.084 %HR	72.4 %HR	0.316 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.125 %HR	91.1 %HR	-1.025 %HR	± 3.0 %HR	± 1.9E+00 %HR