

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.50 μm	-0.15 μm	± 1.6 μm	±	5.8E-02 μm
155.23 μm	155.4 μm	0.17 μm	± 4.7 μm	±	6.2E-02 μm
257.12 μm	257.2 μm	0.08 μm	± 7.7 μm	±	7.0E-02 μm
511.72 μm	511.2 μm	-0.52 μm	± 15.4 μm	±	9.7E-02 μm