

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |             |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|-------------|
| 999.9 RPM                | 1000 RPM         | 0.1 RPM              | ± 20 RPM                    | ±                                  | 5.8E+01 RPM |
| 1499.7 RPM               | 1500 RPM         | 0.3 RPM              | ± 20 RPM                    | ±                                  | 5.8E+01 RPM |
| 1999.9 RPM               | 2000 RPM         | 0.1 RPM              | ± 20 RPM                    | ±                                  | 5.8E+01 RPM |
| 2499.8 RPM               | 2500 RPM         | 0.2 RPM              | ± 20 RPM                    | ±                                  | 5.8E+01 RPM |
| 2999.6 RPM               | 3000 RPM         | 0.4 RPM              | ± 20 RPM                    | ±                                  | 5.8E+01 RPM |