

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
# 0	0.00000	0.002	± 2.00	± 2.1E-01 mm					
0 a 0.012	0.01108	0.920	± 2.00	± 4.3E-01 mm					
0 a 0.37	0.36911	0.890	± 2.00	± 4.3E-01 mm					
0 a 1.07	1.06910	0.900	± 2.00	± 4.8E-01 mm					
0 a 1.98	1.97922	0.780	± 2.00	± 4.8E-01 mm					
0 a 2.75	2.74922	0.780	± 2.00	± 4.8E-01 mm					
0 a 4	3.99934	0.660	± 2.00	± 4.8E-01 mm					
0 a 4.72	4.71935	0.650	± 2.00	± 4.8E-01 mm					
0 a 5.64	5.63942	0.580	± 2.00	± 3.6E-01 mm					
0 a 6.87	6.86948	0.520	± 2.00	± 3.6E-01 mm					
0 a 7.96	7.95950	0.500	± 2.00	± 3.6E-01 mm					
0 a 9.15	9.14957	0.430	± 2.00	± 3.6E-01 mm					
0 a 10.39	10.38958	0.420	± 2.00	± 3.6E-01 mm					
0 a 11.61	11.60965	0.350	± 2.00	± 3.6E-01 mm					
0 a 13	12.99969	0.310	± 2.00	± 3.6E-01 mm					
0 a 14.04	14.03971	0.290	± 2.00	± 3.6E-01 mm					
0 a 15.3	15.29980	0.200	± 2.00	± 3.6E-01 mm					
0 a 16.53	16.52980	0.200	± 2.00	± 3.6E-01 mm					
0 a 17.79	17.78987	0.130	± 2.00	± 3.6E-01 mm					
0 a 18.98	18.97992	0.080	± 2.00	± 3.6E-01 mm					
0 a 20.29	20.29003	-0.030	± 2.00	± 5.5E-01 mm					
0 a 21.25	21.25000	0.000	± 2.00	± 5.5E-01 mm					
0 a 21.39	21.39007	-0.070	± 2.00	± 5.5E-01 mm					
0 a 22.62	22.62012	-0.120	± 2.00	± 5.5E-01 mm					
0 a 24.05	24.05016	-0.160	± 2.00	± 5.5E-01 mm					
0 a 25.07	25.07017	-0.170	± 2.00	± 5.5E-01 mm					
0 a 26.26	26.26026	-0.260	± 2.00	± 5.5E-01 mm					
0 a 27.54	27.54028	-0.280	± 2.00	± 5.5E-01 mm					
0 a 28.73	28.73032	-0.320	± 2.00	± 5.5E-01 mm					
0 a 29.98	29.98039	-0.390	± 2.00	± 5.5E-01 mm					