

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
# 0	-0.00001	0.005	± 0.2	± 2.9E-01 in					
0 a 0.04	0.04012	-0.001	± 0.016	± 2.9E-01 in					
0 a 1.2	1.20016	-0.002	± 0.4	± 2.9E-01 in					
0 a 3.5	3.50001	0.000	± 0.9	± 2.9E-01 in					
0 a 6.49	6.49027	-0.003	± 1.4	± 2.9E-01 in					
0 a 9.01	9.01019	-0.002	± 2	± 2.9E-01 in					
0 a 12.51	12.51040	-0.005	± 2.7	± 2.9E-01 in					
0 a 15.49	15.49032	-0.004	± 3.2	± 2.9E-01 in					
0 a 18.52	18.52043	-0.005	± 3.9	± 2.9E-01 in					
0 a 22.53	22.53046	-0.006	± 4.7	± 2.9E-01 in					
0 a 26.12	26.12039	-0.005	± 5.4	± 2.9E-01 in					
0 a 30.03	30.03046	-0.006	± 6.2	± 2.9E-01 in					
0 a 34.1	34.10031	-0.004	± 7	± 2.9E-01 in					
0 a 38.1	38.10040	-0.005	± 7.8	± 2.9E-01 in					
0 a 42.05	42.05036	-0.004	± 8.6	± 2.9E-01 in					
0 a 46.06	46.06029	-0.004	± 9.4	± 2.9E-01 in					
0 a 50.2	50.20042	-0.005	± 10.2	± 2.9E-01 in					
0 a 54.24	54.24027	-0.003	± 11	± 2.9E-01 in					
0 a 58.35	58.35036	-0.004	± 11.8	± 2.9E-01 in					
0 a 62.27	62.27037	-0.004	± 12.6	± 2.9E-01 in					
0 a 66.58	66.58058	-0.007	± 13.5	± 2.9E-01 in					
0 a 69.73	69.73039	-0.005	± 14.1	± 2.9E-01 in					
0 a 70.19	70.19060	-0.007	± 14.2	± 2.9E-01 in					
0 a 74.21	74.21062	-0.007	± 15	± 2.9E-01 in					
0 a 78.89	78.89058	-0.007	± 15.9	± 2.9E-01 in					
0 a 82.25	82.25050	-0.006	± 16.6	± 2.9E-01 in					
0 a 86.17	86.17067	-0.008	± 17.4	± 2.9E-01 in					
0 a 90.34	90.34059	-0.007	± 18.2	± 2.9E-01 in					
0 a 94.27	94.27060	-0.007	± 19	± 2.9E-01 in					
0 a 98.36	98.36072	-0.009	± 19.8	± 2.9E-01 in					
0 a 99.18	99.18076	-0.0091	± 20	± 2.9E-01 in					