

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
# 0	-0.000001	0.002	± 2.00	± 2.1E-01 mm					
0 a 0.012	0.01110	0.900	± 2.00	± 4.3E-01 mm					
0 a 0.37	0.36912	0.880	± 2.00	± 4.3E-01 mm					
0 a 1.07	1.06909	0.910	± 2.00	± 4.8E-01 mm					
0 a 1.98	1.97920	0.800	± 2.00	± 4.8E-01 mm					
0 a 2.75	2.74923	0.770	± 2.00	± 4.8E-01 mm					
0 a 4	3.99933	0.670	± 2.00	± 4.8E-01 mm					
0 a 4.72	4.71935	0.650	± 2.00	± 4.8E-01 mm					
0 a 5.64	5.63943	0.570	± 2.00	± 3.6E-01 mm					
0 a 6.87	6.86950	0.500	± 2.00	± 3.6E-01 mm					
0 a 7.96	7.95950	0.500	± 2.00	± 3.6E-01 mm					
0 a 9.15	9.14957	0.430	± 2.00	± 3.6E-01 mm					
0 a 10.39	10.38959	0.410	± 2.00	± 3.6E-01 mm					
0 a 11.61	11.60964	0.360	± 2.00	± 3.6E-01 mm					
0 a 13	12.99971	0.290	± 2.00	± 3.6E-01 mm					
0 a 14.04	14.03970	0.300	± 2.00	± 3.6E-01 mm					
0 a 15.3	15.29978	0.220	± 2.00	± 3.6E-01 mm					
0 a 16.53	16.52977	0.230	± 2.00	± 3.6E-01 mm					
0 a 17.79	17.78987	0.130	± 2.00	± 3.6E-01 mm					
0 a 18.98	18.97991	0.090	± 2.00	± 3.6E-01 mm					
0 a 20.29	20.29003	-0.030	± 2.00	± 5.5E-01 mm					
0 a 21.25	21.25001	-0.010	± 2.00	± 5.5E-01 mm					
0 a 21.39	21.39006	-0.060	± 2.00	± 5.5E-01 mm					
0 a 22.62	22.62012	-0.120	± 2.00	± 5.5E-01 mm					
0 a 24.05	24.05015	-0.150	± 2.00	± 5.5E-01 mm					
0 a 25.07	25.07017	-0.170	± 2.00	± 5.5E-01 mm					
0 a 26.26	26.26026	-0.260	± 2.00	± 5.5E-01 mm					
0 a 27.54	27.54025	-0.250	± 2.00	± 5.5E-01 mm					
0 a 28.73	28.73031	-0.310	± 2.00	± 5.5E-01 mm					
0 a 29.98	29.98040	-0.400	± 2.00	± 5.5E-01 mm					