

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09940	0.600	± 0.6	±	4.3E-01 mm						
0 a 0.5	0.49936	0.640	± 0.8	±	4.3E-01 mm						
0 a 1	0.99948	0.520	± 1	±	4.3E-01 mm						
0 a 2	1.99951	0.490	± 1.4	±	4.8E-01 mm						
0 a 3	2.99953	0.470	± 1.8	±	4.8E-01 mm						
0 a 4	3.99951	0.490	± 2.2	±	4.8E-01 mm						
0 a 5	4.99960	0.400	± 2.6	±	4.8E-01 mm						
0 a 6	5.99969	0.310	± 3	±	3.6E-01 mm						
0 a 7	6.99968	0.320	± 3.4	±	3.6E-01 mm						
0 a 8	7.99997	0.030	± 3.8	±	3.6E-01 mm						