

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.30 μm	23.03 μm	-0.27 μm	± 2.5 μm	±	6.2E-01 μm
51.60 μm	51.6 μm	0.03 μm	± 2.5 μm	±	5.9E-01 μm
124.60 μm	124 μm	-0.60 μm	± 2.5 μm	±	5.8E-01 μm
250.30 μm	249 μm	-1.30 μm	± 2.5 μm	±	8.2E-01 μm
486.00 μm	484 μm	-2.00 μm	± 2.5 μm	±	5.8E-01 μm
963.00 μm	961 μm	-2.00 μm	± 2.5 μm	±	6.0E-01 μm