

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
900 RPM	900 RPM	0 RPM	± 20 RPM	±	5.8E-01 RPM
1800 RPM	1800 RPM	0 RPM	± 20 RPM	±	5.8E-01 RPM
3602 RPM	3600 RPM	-2 RPM	± 20 RPM	±	5.8E-01 RPM
4503 RPM	4500 RPM	-3 RPM	± 20 RPM	±	5.8E+01 RPM