

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.130 °C	-0.8 °C	-0.930 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	9.766 °C	8.9 °C	-0.866 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.125 °C	19.3 °C	-0.825 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.365 °C	48.8 °C	-0.565 °C	± 1.0 °C	± 4.0E-01 °C