

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.130 °C	-0.5 °C	-0.630 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	9.766 °C	9.1 °C	-0.666 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.125 °C	19.4 °C	-0.725 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.365 °C	48.2 °C	-1.165 °C	± 1.0 °C	± 4.0E-01 °C