

Certificado No. (Certificate #):		64232		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-12-10		Marca (Brand):		STARRETT	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		0.5000 in	0.5000 in	0.0000 in	± 0.0012 in	±	3.4E-04 in
Exterior / (Exterior)		1.0000 in	1.0000 in	0.0000 in	± 0.0012 in	±	3.4E-04 in
Exterior / (Exterior)		3.0000 in	3.0000 in	0.0000 in	± 0.0012 in	±	3.4E-04 in
Exterior / (Exterior)		4.0000 in	4.0000 in	0.0000 in	± 0.0012 in	±	3.4E-04 in
Exterior / (Exterior)		6.0000 in	6.0000 in	0.0000 in	± 0.0012 in	±	3.4E-04 in
Interior / (Interior)		1.0000 in	1.0000 in	0.0000 in	± 0.0012 in	±	3.7E-04 in
Interior / (Interior)		3.0000 in	3.0000 in	0.0000 in	± 0.0012 in	±	3.7E-04 in
Profundidad / (Depth)		2.0000 in	2.0000 in	0.0000 in	± 0.0012 in	±	3.3E-04 in
Profundidad / (Depth)		4.0000 in	4.0000 in	0.0000 in	± 0.0012 in	±	3.3E-04 in