

| Certificado No. (Certificate #):         |  | 64233             |           | Equipo (Instrument): |                   | Calibrador Pie de Rey       |            |
|------------------------------------------|--|-------------------|-----------|----------------------|-------------------|-----------------------------|------------|
| Fecha de Calibración (Calibration Date): |  | 2025-12-10        |           | Marca (Brand):       |                   | Mitutoyo                    |            |
| Tipo de Medida (Measurement Type)        |  | Patrón (Standard) | UBP (UUT) | Error (Error)        | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
| Exterior / (Exterior)                    |  | 0.500 in          | 0.500 in  | 0.000 in             | ± 0.002 in        | ±                           | 1.0E-03 in |
| Exterior / (Exterior)                    |  | 1.000 in          | 1.000 in  | 0.000 in             | ± 0.002 in        | ±                           | 1.0E-03 in |
| Exterior / (Exterior)                    |  | 3.000 in          | 3.000 in  | 0.000 in             | ± 0.002 in        | ±                           | 1.0E-03 in |
| Exterior / (Exterior)                    |  | 4.000 in          | 3.999 in  | -0.001 in            | ± 0.003 in        | ±                           | 1.0E-03 in |
| Exterior / (Exterior)                    |  | 6.000 in          | 5.999 in  | -0.001 in            | ± 0.003 in        | ±                           | 1.0E-03 in |
| Interior / (Interior)                    |  | 1.000 in          | 0.998 in  | -0.002 in            | ± 0.002 in        | ±                           | 6.7E-04 in |
| Interior / (Interior)                    |  | 2.000 in          | 1.998 in  | -0.002 in            | ± 0.002 in        | ±                           | 9.5E-04 in |