

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1496 RPM	1500 RPM	4 RPM	± 20 RPM	±	5.8E+01 RPM
1997 RPM	2000 RPM	3 RPM	± 20 RPM	±	5.8E+01 RPM
2499 RPM	2500 RPM	1 RPM	± 20 RPM	±	5.8E+01 RPM
2997 RPM	3000 RPM	3 RPM	± 20 RPM	±	5.8E+01 RPM