

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04979	0.210	± 0.6	± 4.3E-01 mm					
0 a 0.1	0.09964	0.360	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14969	0.310	± 0.6	± 4.3E-01 mm					
0 a 1	0.99954	0.460	± 1	± 4.3E-01 mm					
0 a 2	1.99963	0.370	± 1.4	± 4.8E-01 mm					
0 a 3	2.99948	0.520	± 1.8	± 4.8E-01 mm					
0 a 4	3.99963	0.370	± 2.2	± 4.8E-01 mm					
0 a 5	4.99956	0.440	± 2.6	± 4.8E-01 mm					
0 a 6	5.99959	0.410	± 3	± 3.6E-01 mm					
0 a 7	6.99967	0.330	± 3.4	± 3.6E-01 mm					
0 a 8	7.99966	0.340	± 3.8	± 3.6E-01 mm					
0 a 9	8.99971	0.290	± 4.2	± 3.6E-01 mm					
0 a 10	9.99972	0.280	± 4.6	± 3.6E-01 mm					
0 a 11	10.99981	0.190	± 5	± 3.6E-01 mm					
0 a 12	12.00001	-0.010	± 5.4	± 3.6E-01 mm					
0 a 13	12.99995	0.050	± 5.8	± 3.6E-01 mm					
0 a 14	13.99995	0.050	± 6.2	± 3.6E-01 mm					
0 a 15	14.99999	0.010	± 6.6	± 3.6E-01 mm					