

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.346 °C	20.4 °C	0.054 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.039 °C	25.6 °C	0.561 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.092 °C	30.4 °C	0.308 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	34.805 °C	35.5 °C	0.695 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.759 °C	40.3 °C	0.541 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	59.497 %HR	57.2 %HR	-2.297 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.151 %HR	69.1 %HR	-1.051 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	81.163 %HR	82.1 %HR	0.937 %HR	± 3.0 %HR	± 1.9E+00 %HR
90 %HR	90.396 %HR	89.7 %HR	-0.696 %HR	± 3.0 %HR	± 1.9E+00 %HR