

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.812 °C	29.8 °C	-0.012 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.842 %HR	27.4 %HR	-1.442 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.272 %HR	46.5 %HR	-2.772 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.069 %HR	69.7 %HR	-0.369 %HR	± 3.0 %HR	± 1.9E+00 %HR