

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.812 °C	30.0 °C	0.188 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.842 %HR	30.4 %HR	1.558 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.272 %HR	48.3 %HR	-0.972 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.069 %HR	68.8 %HR	-1.269 %HR	± 3.0 %HR	± 1.9E+00 %HR