

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.092 °C	30.2 °C	0.108 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.822 %HR	27.8 %HR	-1.022 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.132 %HR	51.4 %HR	0.268 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.233 %HR	67.6 %HR	-2.633 %HR	± 3.0 %HR	± 1.9E+00 %HR