

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.092 °C	30.4 °C	0.308 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.822 %HR	26 %HR	-2.822 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.132 %HR	48 %HR	-3.132 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.233 %HR	73 %HR	2.767 %HR	± 5.0 %HR	± 1.9E+00 %HR