

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
70.91 μm	71 μm	0.42 μm	± 5 μm	±	5.8E-01 μm
247.33 μm	245 μm	-2.00 μm	± 5 μm	±	5.8E-01 μm
1,491.15 μm	1,491 μm	-0.48 μm	± 5 μm	±	2.4E+00 μm