

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.101 °C	15.3 °C	0.199 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.514 °C	30.5 °C	-0.014 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.233 %HR	49.6 %HR	-0.633 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.703 %HR	71.9 %HR	1.197 %HR	± 3.0 %HR	± 1.9E+00 %HR