

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 15 °C             | 15.101 °C         | 15.5 °C   | 0.399 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.514 °C         | 30.9 °C   | 0.386 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 50 %HR            | 50.233 %HR        | 49.8 %HR  | -0.433 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.703 %HR        | 72.7 %HR  | 1.997 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |