

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.101 °C	15.1 °C	-0.001 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.514 °C	30.8 °C	0.286 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.233 %HR	49.7 %HR	-0.533 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.703 %HR	72.8 %HR	2.097 %HR	± 3.0 %HR	± 1.9E+00 %HR