

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
0.100 in	0.099 in	-0.0010 in	± 0.002 in	± 5.8E-04 in
0.125 in	0.124 in	-0.0010 in	± 0.002 in	± 5.8E-04 in
0.200 in	0.199 in	-0.0010 in	± 0.002 in	± 5.9E-04 in
0.225 in	0.224 in	-0.0010 in	± 0.002 in	± 6.0E-04 in
0.250 in	0.249 in	-0.0010 in	± 0.002 in	± 6.0E-04 in
0.300 in	0.299 in	-0.0010 in	± 0.002 in	± 6.1E-04 in
0.500 in	0.499 in	-0.0010 in	± 0.002 in	± 6.7E-04 in
0.600 in	0.599 in	-0.0010 in	± 0.002 in	± 7.1E-04 in
1.000 in	0.999 in	-0.0010 in	± 0.002 in	± 9.1E-04 in