

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.704 °C	1.9 °C	0.196 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.920 °C	4.1 °C	0.180 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	5.786 °C	5.9 °C	0.114 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.721 °C	7.8 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C