

Certificado No. (Certificate #):		64382		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-12-15		Marca (Brand):		Mitutoyo	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		5.00 mm	5.00 mm	0 µm	± 50 µm	± 3.6E+01 µm	
Exterior / (Exterior)		10.00 mm	10.00 mm	0 µm	± 50 µm	± 3.6E+01 µm	
Exterior / (Exterior)		50.00 mm	50.00 mm	0 µm	± 50 µm	± 3.6E+01 µm	
Exterior / (Exterior)		100.00 mm	100.00 mm	0 µm	± 50 µm	± 3.6E+01 µm	
Exterior / (Exterior)		150.00 mm	150.00 mm	0 µm	± 50 µm	± 3.6E+01 µm	
Interior / (Interior)		25.00 mm	25.00 mm	0 µm	± 50 µm	± 3.0E+01 µm	
Interior / (Interior)		50.00 mm	50.00 mm	0 µm	± 50 µm	± 3.0E+01 µm	
Profundidad / (Depth)		20.00 mm	20.00 mm	0 µm	± 50 µm	± 3.1E+01 µm	
Profundidad / (Depth)		100.00 mm	100.00 mm	0 µm	± 50 µm	± 3.1E+01 µm	