

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
52.65 μm	52.76 μm	0.11 μm	± 1.6 μm	±	5.8E-02 μm
155.23 μm	155.4 μm	0.17 μm	± 4.7 μm	±	6.2E-02 μm
257.12 μm	257.2 μm	0.08 μm	± 7.7 μm	±	7.0E-02 μm
511.72 μm	511.8 μm	0.08 μm	± 15.4 μm	±	9.7E-02 μm