

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49973	0.270	± 0.3	±	4.3E-01 mm						
0 a 1	0.99962	0.380	± 0.4	±	4.3E-01 mm						
0 a 1.5	1.49957	0.430	± 0.5	±	4.8E-01 mm						
0 a 2	1.99951	0.490	± 0.6	±	4.8E-01 mm						
0 a 2.5	2.49948	0.520	± 0.7	±	4.8E-01 mm						
0 a 3	2.99942	0.580	± 0.8	±	4.8E-01 mm						
0 a 3.5	3.49941	0.590	± 0.9	±	4.8E-01 mm						
0 a 4	3.99932	0.680	± 1	±	4.8E-01 mm						
0 a 4.5	4.49925	0.750	± 1.1	±	4.8E-01 mm						
0 a 5	4.99921	0.790	± 1.2	±	4.8E-01 mm						