

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.03	0.02998	0.020	± 0.2	±	4.3E-01 mm						
0 a 0.06	0.05991	0.090	± 0.2	±	4.3E-01 mm						
0 a 0.09	0.08989	0.110	± 0.2	±	4.3E-01 mm						
0 a 0.12	0.11991	0.090	± 0.2	±	4.3E-01 mm						
0 a 0.15	0.14990	0.100	± 0.2	±	4.3E-01 mm						
0 a 0.18	0.17990	0.100	± 0.2	±	4.3E-01 mm						
0 a 0.21	0.20987	0.130	± 0.2	±	4.3E-01 mm						
0 a 0.24	0.23989	0.110	± 0.2	±	4.3E-01 mm						
0 a 0.27	0.26990	0.100	± 0.2	±	4.3E-01 mm						
0 a 0.3	0.29992	0.080	± 0.2	±	4.3E-01 mm						