

Certificado No. (Certificate #):		64461	Equipo (Instrument): INDICADOR DE DIAL		
Fecha de Calibración (Calibration Date):		2025-12-17	Marca (Brand): Mitutoyo		
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm	± 15 µm	±	5.8E+00 µm
1.50 mm	1.50 mm	0 µm	± 15 µm	±	5.8E+00 µm
2.00 mm	2.00 mm	0 µm	± 15 µm	±	5.8E+00 µm
3.00 mm	3.00 mm	0 µm	± 15 µm	±	5.8E+00 µm
4.00 mm	4.00 mm	0 µm	± 15 µm	±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm	± 15 µm	±	5.8E+00 µm
6.00 mm	6.00 mm	0 µm	± 15 µm	±	5.8E+00 µm
8.00 mm	8.00 mm	0 µm	± 15 µm	±	5.8E+00 µm
10.00 mm	10.00 mm	0 µm	± 15 µm	±	5.8E+00 µm