

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.99 ml	1 ml	0.0080 ml	± 0.1 ml	±	3.5E-01 µl
10.00 ml	10 ml	0.0005 ml	± 0.2 ml	±	3.5E-01 µl
100.07 ml	100 ml	-0.0662 ml	± 1.0 ml	±	3.5E-01 µl
500.06 ml	500 ml	-0.0637 ml	± 5.0 ml	±	3.5E-01 µl
999.94 ml	1000 ml	0.0591 ml	± 10 ml	±	3.5E-01 µl