

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
0.99 ml	1 ml	0.0095 ml	± 0.1 ml	±	3.5E-01 µl
10.00 ml	10 ml	0.0035 ml	± 0.2 ml	±	3.5E-01 µl
100.06 ml	100 ml	-0.0631 ml	± 1.0 ml	±	3.5E-01 µl
500.09 ml	500 ml	-0.0867 ml	± 5.0 ml	±	3.5E-01 µl
999.90 ml	1000 ml	0.0982 ml	± 10 ml	±	3.5E-01 µl