

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
100.13 µl	100 µl	-0.1318 µl	± 0.8 µl	±	5.1E-02 µl
499.71 µl	500 µl	0.2932 µl	± 4 µl	±	5.9E-02 µl
999.91 µl	1000 µl	0.0852 µl	± 8 µl	±	7.7E-02 µl