

Certificado No. (Certificate #):64475Equipo (Instrument):Calibrador Pie de Rey

Fecha de Calibración (Calibration Date):2025-06-20Marca (Brand):Mitutoyo

| Tipo de Medida (Measurement Type) | Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |
|-----------------------------------|-------------------|-----------|---------------|-------------------|-----------------------------|
| Exterior / (Exterior)             | 0.1000 in         | 0.0995 in | -0.0005 in    | ± 0.0012 in       | ± 3.5E-04 in                |
| Exterior / (Exterior)             | 0.5000 in         | 0.4998 in | -0.0002 in    | ± 0.0012 in       | ± 3.5E-04 in                |
| Exterior / (Exterior)             | 4.0000 in         | 4.0000 in | 0.0000 in     | ± 0.0012 in       | ± 3.5E-04 in                |
| Exterior / (Exterior)             | 6.0000 in         | 6.0003 in | 0.0003 in     | ± 0.0012 in       | ± 3.5E-04 in                |
| Exterior / (Exterior)             | 7.5000 in         | 7.5005 in | 0.0005 in     | ± 0.0012 in       | ± 3.5E-04 in                |
| Interior / (Interior)             | 3.0000 in         | 3.0000 in | 0.0000 in     | ± 0.0012 in       | ± 3.7E-04 in                |
| Interior / (Interior)             | 5.0000 in         | 5.0005 in | 0.0005 in     | ± 0.0012 in       | ± 3.7E-04 in                |
| Profundidad / (Depth)             | 1.0000 in         | 0.9998 in | -0.0002 in    | ± 0.0012 in       | ± 4.7E-04 in                |
| Profundidad / (Depth)             | 2.0000 in         | 1.9995 in | -0.0005 in    | ± 0.0012 in       | ± 3.3E-04 in                |