

Certificado No. (Certificate #):		64497		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-12-17		Marca (Brand):		Starrett	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		15.00 mm	15.01 mm	10 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		50.00 mm	50.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		100.00 mm	100.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		135.00 mm	135.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		150.00 mm	150.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Interior / (Interior)		20.00 mm	19.97 mm	-30 µm	± 30 µm	± 8.2E+00 µm	
Interior / (Interior)		100.00 mm	99.97 mm	-30 µm	± 30 µm	± 8.2E+00 µm	
Profundidad / (Depth)		30.00 mm	30.00 mm	0 µm	± 30 µm	± 7.1E+00 µm	
Profundidad / (Depth)		100.00 mm	100.00 mm	0 µm	± 30 µm	± 7.1E+00 µm	